



Form GSOP 1-PIN (04/98)

STATE OF CALIFORNIA
Department of General Services - Office of Procurement

PURCHASE ORDER

Page 1

Purchase Order No. Rev. Date
62029 6/ 6/2008

Supplier No. 807628	Solicitation No. 56674	Delivery Date 60 Days ARO	FOB Point Destination	Invoice Terms
<i>S</i> ATASCADERO STATE HOSPITAL <i>h</i> 10333 EL CAMINO REAL <i>i</i> ATASCADERO, CA 93422 <i>p</i> Attn: RECEIVING DOCK		<i>C</i> ATASCADERO STATE HOSPITAL <i>h</i> PO BOX 7002 <i>a</i> ATASCADERO CA 93423 <i>r</i> ACCOUNTS PAYABLE <i>e</i>		
Agency Billing 49050	Agency Purchase Estimate 61201005	Purchase Estimate 66677	Revision 0	
Agency Contact DARLA CARNES		Phone 805-468-3006	Date Received	

DC INSTRUMENTS
2546 RIDGEBROOK PL
THOUSAND OAKS, CA 91362
Attn: DANIEL CHANG

Phone: 805-300-6546

Item No.	Quantity	Unit	Commodity Code	Description	Unit Price	Extension
<p>THE GENERAL PROVISIONS FOR NON-IT COMMODITIES ARE HEREBY INCORPORATED BY REFERENCE. THESE GENERAL PROVISIONS CAN BE OBTAINED BY PHONING (916) 375-4400 OR BY ACCESSING OUR WEBSITE AT:</p> <p>www.documents.dgs.ca.gov/pd/modellang/GPnonIT0407.pdf</p> <p>THE FOLLOWING INFORMATION IS PROVIDED FOR AGENCY USE ONLY:</p> <p>PRIME CONTRACTOR: MB;SB</p> <p>APPLICATION OF DVBE INCENTIVE RESULTED IN AWARD: NO</p>						
1	35	EA	6530-002-0128-1	TABLE MEDICAL (AS DESCRIBED) Treatment table; as described meeting the requirement of the attached specification 6530-07BS-005 of two pages.	1,723.2900	60,315.15
				Brand: HAUSMANN DELUXE TREATMENT TBL Model: 4796		
2	40	EA	1096-000-0194-0	STOOL STEP Step Stool; as described meeting the requirements of the attached specification 6530-07BS-005 of two pages.	29.0000	1,160.00
				Brand: PROFESSIONAL MEDICAL Model: STOOL		
3	40	EA	6530-999-0005-9	CART (FOR MEDICAL USE) (AS DESCRIBED) Rolling Chart Rack; as described meeting the requirements of the attached specification 6530-07BS-005 of two pages.	353.9900	14,159.60
				Brand: WINCO Model: 7161		

Sales and/or use tax to be extra unless noted above

Buyer GUS QUINTERO	Phone 916-375-4499	BOC Number
-----------------------	-----------------------	------------

Department of General Services - Office of Procurement

Page 2

Form GSOP 2-PIN (04/98)

<i>Purchase Order No.</i>	<i>Revision</i>	<i>Date</i>	<i>Supplier No.</i>	<i>Supplier Name</i>
62029		6/ 6/2008	807628	DC INSTRUMENTS

<i>Item No.</i>	<i>Quantity</i>	<i>Unit</i>	<i>Commodity Code</i>	<i>Description</i>	<i>Unit Price</i>	<i>Extension</i>
4	12	EA	6515-478-0100-2	STETHOSCOPE Stethoscope; as described meeting the requirements of the attached specification 6530-07BS-005 of two pages. Brand: <u>3M LITTMANN CLASSIC II S.E.</u> Model: <u>2201</u>	58.5000	702.00
5	34	EA	6650-566-0202-0	MEDICAL INSTRUMENT (AS DESCRIBED) Tunning Forks 128 cps (cycles per second); as described meeting the requirements of the attached specification 6530-07BS-005 of two pages. Brand: <u>MEDICAL PROFESSIONAL</u> Model: <u>C128HZ</u>	8.7500	297.50
6	34	EA	6650-566-0202-0	MEDICAL INSTRUMENT (AS DESCRIBED) Tunning Forks 256 cps (cycles per second); as described meeting the requirements of the attached specification 6530-07BS-005 of two pages. Brand: <u>MEDICAL PROFESSIONAL</u> Model: <u>C256HZ</u>	8.7500	297.50
7	34	EA	6530-999-0005-9	CART (FOR MEDICAL USE) (AS DESCRIBED) Rolling Chart cart 9Cabinet); as described meeting the requirements of the attached specification 6530-07BS-05 of two pages. Brand: <u>PMNI MED</u> Model: <u>260120</u>	564.6200	19,197.08
<u>PO Miscellaneous Charges and Discounts</u>						<u>Dollar Value</u>
DELIVERY CHARGE						2,000.00
<u>Total Value:</u>						<u>98,128.83</u>
<u>FOB DESTINATION:</u>						
For the purpose of this order, only F.O.B. Destination will be accepted.						

Department of General Services - Office of Procurement

Form GSOP 2-PIN (04/98)

Page 3 (Last)

Item No.	Quantity	Unit	Commodity Code	Description	Unit Price	Extension
<p><u>ELECTRICAL</u></p> <p>All electrical and mechanical equipment furnished shall comply with the California Administrative Code; Title 8 (Industrial Safety Orders), Title 24 (State Building Standards Law) and Title 17 (Public Health). All electrical equipment furnished shall be grounded, with any exceptions only as approved in the referenced applicable titles.</p> <p><u>UNPACK AND SET IN PLACE:</u></p> <p>The vendor will be required to unpack and set the material in place as directed, remove all dunnage and instruct agency operating personnel in its use.</p> <p><u>WARRANTY:</u></p> <p>Equipment shall operate satisfactorily and have a minimum warranty period of one year from date of delivery to the State. Vendor shall bear all material, labor and transportation costs for repair of defects and failures occurring within the warranty period.</p> <p><u>MANUALS:</u></p> <p>Each system must come with one operation manual, service manual, and parts list (If applicable). If there are any special tools required for basic maintenance, one shall be provided. (All manuals are to be in English)</p> <p>CHANGE ORDERS:</p> <p>Any Purchase Order resulting from this bid may be amended, modified, or terminated at any time by mutual agreement of the parties in writing. Change orders amending, modifying or terminating the Purchase Order, including any modifications of the compensation payable, may be issued only by the State Procurement Officer. All such change orders shall be in writing and issued only upon written concurrence of the supplier. Termination, as that term is used in this section, does not include termination for default of the supplier.</p> <p>This Purchase order has been registered into the state contact and procurement registration system (https://www.scprs.dgs.ca.gov/). The registration number is 46000608310760.</p>						



STATE OF CALIFORNIA
BID SPECIFICATION
HOSPITAL SUPPLIES

6530-07BS-005



1. SCOPE




This document describes hospital supplies utilized in an examination room by Atascadero State Hospital, State of California.

2. APPLICABLE SPECIFICATIONS / STANDARDS

Specifications and standards referenced in this document in effect on the opening of the Invitation for Bid, form a part of this specification

3. GENERAL

Item #1	<p>Treatment Table TO INCLUDE THE FOLLOWING MINIMUM FEATURES:</p> <ul style="list-style-type: none">• (2) Drawers• Two storage Door sections each with adjustable shelf• 0 degrees - 75 degrees backrest with control lever• Equipped with Paper Dispenser and Cutter Combo• 2" Vinyl upholstery finish• 400 lbs. weight capacity• Fit through a door way 34"Wx79"H <p>REFERENCE BRAND: Hausmann 4796 Deluxe Treatment Table REFERENCE MODEL: 4796 OR EQUIVALENT</p>	
Item #2	<p>Step Stool TO INCLUDE THE FOLLOWING MINIMUM FEATURES:</p> <ul style="list-style-type: none">• Stool has a chrome-plated finish• The top has a skid-resistant top.• The rubber leg-tips also provide skid-resistance.• Weight Capacity: 250 lbs• Height: 9"• Width: 14" x 11" <p>REFERENCE BRAND: ABCO Step Stool REFERENCE MODEL: 371120 OR EQUIVALENT</p>	
Item #3	<p>Rolling Chart Rack (Cabinet) TO INCLUDE THE FOLLOWING MINIMUM FEATURES:</p> <ul style="list-style-type: none">• Fully enclosed top, side and back panels (2 enclosed shelves with locking front panels)• Equipped to hold 16 with spine measurement up to 3"• 4 rolling casters (2 with wheel locks)• Approximate Dimensions (41" high x 29.5" wide) <p>REFERENCE BRAND: Medline Chart Rack REFERENCE MODEL: FHP638016P OR EQUIVALENT</p>	No Picture available

Item #4	<p>Stethoscope TO INCLUDE THE FOLLOWING MINIMUM FEATURES:</p> <ul style="list-style-type: none"> • Two-sided chest-piece that offers high acoustic sensitivity. • Tunable diaphragm on one side augments the traditional bell function of the opposite side. • Single-lumen tubing • 28" tube <p>REFERENCE BRAND: 3M Littmann Classic II S.E. Stethoscope REFERENCE MODEL: 2201 OR EQUIVALENT</p>	
Item #5	<p>TUNING FORKS TO INCLUDE THE FOLLOWING MINIMUM FEATURES:</p> <p>Aluminum alloy tuning forks designed and manufactured to meet the demanding needs of the medical profession.</p> <ul style="list-style-type: none"> • 128 cps (cycles per second). • Non-magnetic, corrosion resistant aluminum alloy construction • long 2" handle to facilitate bone conduction tests <p>REFERENCE BRAND: ADC Tuning Fork 128cps Frequency REFERENCE MODEL: 500128 OR EQUIVALENT</p>	
Item #6	<p>TUNING FORKS TO INCLUDE THE FOLLOWING MINIMUM FEATURES:</p> <p>Aluminum alloy tuning forks designed and manufactured to meet the demanding needs of the medical profession.</p> <ul style="list-style-type: none"> • 256 cps (cycles per second) • Non-magnetic, corrosion resistant aluminum alloy construction • long 2" handle to facilitate bone conduction tests <p>REFERENCE BRAND: ADC Tuning Fork 256cps Frequency REFERENCE MODEL: 500256 OR EQUIVALENT</p>	
Item #7	<p>Rolling Chart Cart (Cabinet) TO INCLUDE THE FOLLOWING MINIMUM FEATURES:</p> <ul style="list-style-type: none"> • Fully enclosed top, side and back panels (2 enclosed shelves) • Equipped to hold 20 binders with spine measurement up to 2.75". • 4 rolling casters (2 with wheel locks) • Equipped with binder dividers • Approximate Dimensions (43.75" high x 34.5" wide x 17" deep) <p>REFERENCE BRAND: Omnimed 260120 Beam Standard Cabinet-Style REFERENCE MODEL: 260120 OR EQUIVALENT</p>	